Certificate Number Report Reference 20180816-E486483 E486483-20170127 2018-AUGUST-16

Issue Date

Issued to:

mPower Electronics

3046 Scott Blvd

Santa Clara CA 95054

This is to certify that representative samples of

GAS AND VAPOR DETECTION EQUIPMENT FOR USE IN

HAZARDOUS LOCATIONS

Single gas detector, Model MP100, MP101, MP102 and MP103, intrinsically safe when used with mPower battery

Part No. 1.17.02.0002 or EVE Part No. ER14505.

Have been investigated by UL in accordance with the

Standard(s) indicated on this Certificate.

Standard(s) for Safety:

UL 913, Intrinsically Safe Apparatus and Associated

Apparatus for Use in Class I, II, and III, Division 1,

Hazardous (Classified) Locations

UL 508 and CSA C22.2 No. 14, Standard for Industrial

Control Equipment

CSA-C22.2 No. 157-92, Intrinsically Safe and Nonincendive Equipment for Use in Hazardous Locations

Additional Information:

See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.

Bruce Mahrenholz Director North American Continuing Brown

DLITC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized scenses of UL. For questions, please contact a local UL Customer Service Representation at http://link.com/physiolate/firestimes/



Certificate Number Report Reference 20190607-E486483 E486483-20190601

Issue Date

2019-JUNE-07

Issued to:

mPower Electronics 3046 Scott Blvd Santa Clara CA 95054

This certificate confirms that representative samples of GAS AND VAPOR DETECTION EQUIPMENT FOR USE IN HAZARDOUS LOCATIONS

Class I, Groups A, B, C and D Hazardous Locations.

Multi gas detector, Model MP400, optionally followed by P, S or H; intrinsically safe when used with mPower battery

Part No. 12.01.09.0001 or 12.01.09.0003.

Have been investigated by UL in accordance with the

Standard(s) indicated on this Certificate.

Standard(s) for Safety:

UL 913 - Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1,

Hazardous (Classified) Locations

UL 508 - Standard for Industrial Control Equipment

Additional Information:

See the UL Online Certifications Directory at

https://ig.ulprospector.com for additional information.

This Certificate of Compliance does not provide authorization to apply the UL Mark. Only the UL Follow-Up Services Procedure provides authorization to apply the UL Mark.

Only those products bearing the UL Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.

Look for the UL Certification Mark on the product.

Bruce Mahrenholz, Director North American Certification Program

NUTTO

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized 8 censes of UL. For questions, please contact a local LL Customer Service Representative at https://discontinent/



Certificate Number Report Reference 20190607-E486483 E486483-20190531

Issue Date

2019-June-07

Issued to:

mPower Electronics

3046 Scott Blvd

Santa Clara CA 95054

This certificate confirms that representative samples of GAS AND VAPOR DETECTION EQUIPMENT FOR USE IN

HAZARDOUS LOCATIONS

USC, CNC – Gas and Vapor Detection Equipment for Use in Class I, Division 1, Group A, B, C and D Hazardous

Locations

NEO Portable VOC Gas Detector, Part No. MP18, followed by 0, 1, 2, 3, 4, 5, 6, 7, 8 or 9, intrinsically safe when used

with battery pack part no. M011-3002-000.

Have been investigated by UL in accordance with the

Standard(s) indicated on this Certificate.

Standard(s) for Safety: Additional Information: SEE ADDENDUM PAGE FOR Standard(s)

See the UL Online Certifications Directory at

https://iq.ulprospector.com for additional information.

This Certificate of Compliance does not provide authorization to apply the UL Mark.

Only those products bearing the UL Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.

Look for the UL Certification Mark on the product.



ULLLO

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Represents two at https://ul.com/shoutu/focations/



Certificate Number Report Reference 20190607-E486483 E486483-20190531

Issue Date 2019-June-07

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Standard(s)

UL 913, Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations

UL 61010-1, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CAN/CSA C22.2 No. 61010-1, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CAN/CSA C22.2 No. 60079-0, Explosive Atmospheres - Part 0: Equipment - General Requirements

CAN/CSA C22.2 No. 60079-11, Explosive Atmospheres - Part 11: Equipment Protection by Intrinsic Safety "i"



